



HC3: Alert

June 04, 2021

TLP: White

Report: 202106041315

Vulnerabilities Reported by MesaLabs for AmegaView

Executive Summary

MesaLabs' AmegaView is a continuous monitoring hardware and software platform for laboratories. MesaLabs has announced the discovery of five vulnerabilities, including two critical command injection vulnerabilities, that could allow remote code execution or allow access to devices running AmegaView Versions 3.0 and prior.

Impact to HPH Sector

Although there are currently no public exploits actively targeting these vulnerabilities, the MesaLabs AmegaView continuous monitoring system is used by hospital laboratories, forensics labs, blood banks, IVF labs, pharmacies, and biotech firms. HC3 recommends affected HPH organizations locate vulnerable products behind firewalls, isolate them from the network, and ensure they are only accessible to the Internet using a secure and updated Virtual Private Network (VPN). MesaLabs has announced that it will not release patches to correct the vulnerabilities, although users can upgrade to newer Viewpoint software compatible with AmegaView hardware that is not affected by these vulnerabilities.

- CVE-2021-27447 – CVSS 10/10 – Flaw due to improper neutralization of special elements used in a command, which could allow an attacker to execute arbitrary code.
- CVE-2021-27449 – CVSS 9.9/10 – Flaw due to improper neutralization of special elements used in a command, which could allow an attacker to execute commands in the web server.
- CVE-2021-27445 – CVSS 7.8/10 – Insecure file permissions which could be exploited to elevate privileges on the device.
- CVE-2021-27451 – CVSS 7.3/10 – Improper authentication due to passcodes being generated by an easily reversible algorithm, which could allow an attacker to gain access to the device.
- CVE-2021-27453 – CVSS 7.3/10 – Authentication bypass issue that could allow an attacker to gain access to the web application.

References

Critical Vulnerabilities identified in MesaLabs Laboratory Temperature Monitoring System

https://www.hipaajournal.com/critical-vulnerabilities-identified-in-mesalabs-laboratory-temperature-monitoring-system/amp/?_twitter_impression=true

ICS Advisory (ICSA-21-147-03)

<https://us-cert.cisa.gov/ics/advisories/icsa-21-147-03>

Contact Information

If you have any additional questions, please contact us at HC3@hhs.gov.

We want to know how satisfied you are with our products. Your answers will be anonymous, and we will use the responses to improve all our future updates, features, and new products. [Share Your Feedback](#)